


# **Volatile Organic Compound Analysis Results for Samples Collected in Nuiqsut, Alaska**

Sample Location: Nuiqsut Ambient Air Quality Monitoring Station

Date Sample Collected: 3/10/2022

Analysis Conducted by: Atmospheric Analysis & Consulting, Inc.

Analysis Method: EPA Method TO-12/PAMS Protocol by GC/MS/FID



# Atmospheric Analysis & Consulting, Inc.


## Laboratory Analysis Report

**CLIENT** : SLR International Corporation  
**PROJECT NO** : 220531  
**MATRIX** : AIR  
**UNITS** : ppbC

**DATE RECEIVED** : 03/14/2022  
**DATE REPORTED** : 03/15/2022

### HYDROCARBONS (C2-C12) SPECIATED

<i>Client ID</i>	NUI - unlined			Sample Reporting Limit (SRL) (MRLxDFs)	Method Reporting Limit (MRL)
<i>AAC ID</i>	220531-28822				
<i>Date Sampled</i>	03/10/2022				
<i>Date Analyzed</i>	03/14/2022				
<i>Can Dilution Factor</i>	1.55				
	Result	Qualifier	Analysis DF		
Ethylene	2.57		1	1.55	1.00
Acetylene	9.01		1	1.55	1.00
Ethane	5.20		1	1.55	1.00
Propylene	<SRL	U	1	1.55	1.00
Propane	3.33		1	1.55	1.00
Isobutane	<SRL	U	1	1.55	1.00
1-Butene	<SRL	U	1	1.55	1.00
n-Butane	<SRL	U	1	1.55	1.00
trans-2-Butene	<SRL	U	1	1.55	1.00
cis-2-Butene	<SRL	U	1	1.55	1.00
Isopentane	<SRL	U	1	1.55	1.00
1-Pentene	<SRL	U	1	1.55	1.00
n-Pentane	<SRL	U	1	1.55	1.00
Isoprene	<SRL	U	1	1.55	1.00
trans-2-Pentene	<SRL	U	1	1.55	1.00
cis-2-Pentene	<SRL	U	1	1.55	1.00
2,2-Dimethylbutane	<SRL	U	1	1.55	1.00
Cyclopentane	<SRL	U	1	1.55	1.00
2,3-Dimethylbutane	<SRL	U	1	1.55	1.00
2-Methylpentane	<SRL	U	1	1.55	1.00
3-Methylpentane	<SRL	U	1	1.55	1.00
1-Hexene	<SRL	U	1	1.55	1.00
n-Hexane	<SRL	U	1	1.55	1.00
Methylcyclopentane	<SRL	U	1	1.55	1.00
2,4-Dimethylpentane	<SRL	U	1	1.55	1.00
Benzene	<SRL	U	1	1.55	1.00
Cyclohexane	<SRL	U	1	1.55	1.00
2-Methylhexane	<SRL	U	1	1.55	1.00
2,3-Dimethylpentane	<SRL	U	1	1.55	1.00
3-Methylhexane	<SRL	U	1	1.55	1.00
2,2,4-Trimethylpentane	<SRL	U	1	1.55	1.00
n-Heptane	<SRL	U	1	1.55	1.00
Methylcyclohexane	<SRL	U	1	1.55	1.00
2,3,4-Trimethylpentane	<SRL	U	1	1.55	1.00



# Atmospheric Analysis & Consulting, Inc

## Laboratory Analysis Report


**CLIENT** : SLR International Corporation  
**PROJECT NO** : 220531  
**MATRIX** : AIR  
**UNITS** : ppbC

**DATE RECEIVED** : 03/14/2022  
**DATE REPORTED** : 03/15/2022

### HYDROCARBONS (C2-C12) SPECIATED

<i>Client ID</i>	NUI - unlined			Sample Reporting Limit (SRL) (MRLxDFs)	Method Reporting Limit (MRL)
<i>AAC ID</i>	220531-28822				
<i>Date Sampled</i>	03/10/2022				
<i>Date Analyzed</i>	03/14/2022				
<i>Can Dilution Factor</i>	1.55				
	Result	Qualifier	Analysis DF		
Toluene	3.61		1	1.55	1.00
2-Methylheptane	<SRL	U	1	1.55	1.00
3-Methylheptane	<SRL	U	1	1.55	1.00
n-Octane	<SRL	U	1	1.55	1.00
Ethylbenzene	2.51		1	1.55	1.00
m/p-Xylenes	4.12		1	1.55	1.00
Styrene	<SRL	U	1	1.55	1.00
o-Xylene	1.64		1	1.55	1.00
Nonane	<SRL	U	1	1.55	1.00
Isopropylbenzene	<SRL	U	1	1.55	1.00
n-Propylbenzene	<SRL	U	1	1.55	1.00
m-Ethyltoluene	<SRL	U	1	1.55	1.00
p-Ethyltoluene	<SRL	U	1	1.55	1.00
1,3,5-Trimethylbenzene	<SRL	U	1	1.55	1.00
o-Ethyltoluene	<SRL	U	1	1.55	1.00
1,2,4-Trimethylbenzene	<SRL	U	1	1.55	1.00
n-Decane	<SRL	U	1	1.55	1.00
1,2,3-Trimethylbenzene	<SRL	U	1	1.55	1.00
m-Diethylbenzene	<SRL	U	1	1.55	1.00
p-Diethylbenzene	<SRL	U	1	1.55	1.00
n-Undecane	<SRL	U	1	1.55	1.00
n-Dodecane	<SRL	U	1	1.55	1.00

U - Compound was analyzed for, but was not detected at or above the SRL.



# Atmospheric Analysis & Consulting, Inc.

---

## Quality Control/Quality Assurance Report PAMS Calibration Verification Analysis

Initial Calibration Date : 02/11/2022  
Standard ID : MS1-020922-01

Instrument ID : MS01  
Analysis Date : 03/14/2022  
Analyst : RB

### Continuing Calibration Verification

Propane	xRF	Daily RF	RPD*
	698	691	1.04


\* Must be <10%

### Laboratory Control Spike (LCS) / Laboratory Control Spike Duplicate (LCSD)

Propane	Sample Conc. (ppbC)	Spike Added (ppbC)	Recovery (ppbC)		% Recovery**		RPD***
			LCS	LCSD	LCS	LCSD	
	0.00	4.24	4.19	4.18	98.8	98.6	0.24

\*\* Must be 80-120%

\*\*\* Must be <25%



# Atmospheric Analysis & Consulting, Inc.

## Quality Control/Quality Assurance Report


### PAMS Method Blank Analysis

Matrix : Air  
Units : ppbC

Instrument ID : MS01  
Analysis Date : 03/14/2022  
Analyst : RB

Analyte	Result	PQL
Ethylene	<PQL	1.0
Acetylene	<PQL	1.0
Ethane	<PQL	1.0
Propylene	<PQL	1.0
Propane	<PQL	1.0
Isobutane	<PQL	1.0
1-Butene	<PQL	1.0
n-Butane	<PQL	1.0
trans-2-Butene	<PQL	1.0
cis-2-Butene	<PQL	1.0
Isopentane	<PQL	1.0
1-Pentene	<PQL	1.0
n-Pentane	<PQL	1.0
Isoprene	<PQL	1.0
trans-2-Pentene	<PQL	1.0
cis-2-Pentene	<PQL	1.0
2,2-Dimethylbutane	<PQL	1.0
Cyclopentane	<PQL	1.0
2,3-Dimethylbutane	<PQL	1.0
2-Methylpentane	<PQL	1.0
3-Methylpentane	<PQL	1.0
1-Hexene	<PQL	1.0
n-Hexane	<PQL	1.0
Methylcyclopentane	<PQL	1.0
2,4-Dimethylpentane	<PQL	1.0
Benzene	<PQL	1.0
Cyclohexane	<PQL	1.0
2-Methylhexane	<PQL	1.0
2,3-Dimethylpentane	<PQL	1.0
3-Methylhexane	<PQL	1.0
2,2,4-Trimethylpentane	<PQL	1.0
n-Heptane	<PQL	1.0
Methylcyclohexane	<PQL	1.0
2,3,4-Trimethylpentane	<PQL	1.0

Analyte	Result	PQL
Toluene	<PQL	1.0
2-Methylheptane	<PQL	1.0
3-Methylheptane	<PQL	1.0
n-Octane	<PQL	1.0
Ethylbenzene	<PQL	1.0
m/p-Xylenes	<PQL	1.0
Styrene	<PQL	1.0
o-Xylene	<PQL	1.0
Nonane	<PQL	1.0
Isopropylbenzene	<PQL	1.0
n-Propylbenzene	<PQL	1.0
m-Ethyltoluene	<PQL	1.0
p-Ethyltoluene	<PQL	1.0
1,3,5-Trimethylbenzene	<PQL	1.0
o-Ethyltoluene	<PQL	1.0
1,2,4-Trimethylbenzene	<PQL	1.0
n-Decane	<PQL	1.0
1,2,3-Trimethylbenzene	<PQL	1.0
m-Diethylbenzene	<PQL	1.0
p-Diethylbenzene	<PQL	1.0
n-Undecane	<PQL	1.0
n-Dodecane	<PQL	1.0
TNMHC (ppbC)	<PQL	20



# Atmospheric Analysis & Consulting, Inc.

## Quality Control/Quality Assurance Report PAMS Duplicate Analysis

AAC ID : 220507-28820  
 Matrix : Air  
 Units : ppbC

Instrument ID : MS01  
 Analysis Date : 03/14/2022  
 Analyst : RB

Analyte	Sample Analysis	Sample Duplicate	RPD
Ethylene	<PQL	<PQL	NA
Acetylene	<PQL	<PQL	NA
Ethane	4.69	5.25	11
Propylene	<PQL	<PQL	NA
Propane	2.79	2.96	5.9
Isobutane	<PQL	<PQL	NA
1-Butene	<PQL	<PQL	NA
n-Butane	<PQL	<PQL	NA
trans-2-Butene	<PQL	<PQL	NA
cis-2-Butene	<PQL	<PQL	NA
Isopentane	<PQL	<PQL	NA
1-Pentene	<PQL	<PQL	NA
n-Pentane	<PQL	<PQL	NA
Isoprene	<PQL	<PQL	NA
trans-2-Pentene	<PQL	<PQL	NA
cis-2-Pentene	<PQL	<PQL	NA
2,2-Dimethylbutane	<PQL	<PQL	NA
Cyclopentane	<PQL	<PQL	NA
2,3-Dimethylbutane	<PQL	<PQL	NA
2-Methylpentane	<PQL	<PQL	NA
3-Methylpentane	<PQL	<PQL	NA
1-Hexene	<PQL	<PQL	NA
n-Hexane	<PQL	<PQL	NA
Methylcyclopentane	<PQL	<PQL	NA
2,4-Dimethylpentane	<PQL	<PQL	NA
Benzene	<PQL	<PQL	NA
Cyclohexane	<PQL	<PQL	NA
2-Methylhexane	<PQL	<PQL	NA
2,3-Dimethylpentane	<PQL	<PQL	NA
3-Methylhexane	<PQL	<PQL	NA
2,2,4-Trimethylpentane	<PQL	<PQL	NA
n-Heptane	<PQL	<PQL	NA
Methylcyclohexane	<PQL	<PQL	NA
2,3,4-Trimethylpentane	<PQL	<PQL	NA
Toluene	<PQL	<PQL	NA
2-Methylheptane	<PQL	<PQL	NA
3-Methylheptane	<PQL	<PQL	NA
n-Octane	<PQL	<PQL	NA
Ethylbenzene	<PQL	<PQL	NA
m/p-Xylenes	<PQL	<PQL	NA
Styrene	<PQL	<PQL	NA
o-Xylene	<PQL	<PQL	NA
Nonane	<PQL	<PQL	NA
Isopropylbenzene	<PQL	<PQL	NA
n-Propylbenzene	<PQL	<PQL	NA
m-Ethyltoluene	<PQL	<PQL	NA
p-Ethyltoluene	<PQL	<PQL	NA

Analyte	Sample Analysis	Sample Duplicate	RPD
1,3,5-Trimethylbenzene	<PQL	<PQL	NA
o-Ethyltoluene	<PQL	<PQL	NA
1,2,4-Trimethylbenzene	<PQL	<PQL	NA
n-Decane	<PQL	<PQL	NA
1,2,3-Trimethylbenzene	<PQL	<PQL	NA
m-Diethylbenzene	<PQL	<PQL	NA
p-Diethylbenzene	<PQL	<PQL	NA
n-Undecane	<PQL	<PQL	NA
n-Dodecane	<PQL	<PQL	NA
Total PAMS (ppbC)	7.47	8.20	9.3
TNMHC (ppbC)	30.6	34.9	13