


Volatile Organic Compound Analysis Results for Samples Collected in Nuiqsut, Alaska

Sample Location: Nuiqsut Ambient Air Quality Monitoring Station

Date Sample Collected: 3/12/2022

Analysis Conducted by: Atmospheric Analysis & Consulting, Inc.

Analysis Method: EPA Method TO-12/PAMS Protocol by GC/MS/FID



Atmospheric Analysis & Consulting, Inc.


Laboratory Analysis Report

CLIENT : SLR International Corporation
PROJECT NO : 220526 Rev 1
MATRIX : AIR
UNITS : ppb (v/v)

DATE RECEIVED : 03/15/2022
DATE REPORTED : 03/16/2022

HYDROCARBONS (C2-C12) SPECIATED

<i>Client ID</i>	NUI DUP			Sample Reporting Limit (SRL) (MRLxDFs)	Method Reporting Limit (MRL)
<i>AAC ID</i>	220526-28954				
<i>Date Sampled</i>	03/12/2022				
<i>Date Analyzed</i>	03/15/2022				
<i>Can Dilution Factor</i>	1.49				
	Result	Qualifier	Analysis DF		
Ethylene	<SRL	U	1	0.75	0.50
Acetylene	<SRL	U	1	0.75	0.50
Ethane	3.65		1	0.75	0.50
Propylene	<SRL	U	1	0.50	0.33
Propane	1.59		1	0.50	0.33
Isobutane	<SRL	U	1	0.37	0.25
1-Butene	<SRL	U	1	0.37	0.25
n-Butane	0.46		1	0.37	0.25
trans-2-Butene	<SRL	U	1	0.37	0.25
cis-2-Butene	<SRL	U	1	0.37	0.25
Isopentane	<SRL	U	1	0.30	0.20
1-Pentene	<SRL	U	1	0.30	0.20
n-Pentane	<SRL	U	1	0.30	0.20
Isoprene	<SRL	U	1	0.30	0.20
trans-2-Pentene	<SRL	U	1	0.30	0.20
cis-2-Pentene	<SRL	U	1	0.30	0.20
2,2-Dimethylbutane	<SRL	U	1	0.25	0.17
Cyclopentane	<SRL	U	1	0.30	0.20
2,3-Dimethylbutane	<SRL	U	1	0.25	0.17
2-Methylpentane	<SRL	U	1	0.25	0.17
3-Methylpentane	<SRL	U	1	0.25	0.17
1-Hexene	<SRL	U	1	0.25	0.17
n-Hexane	<SRL	U	1	0.25	0.17
Methylcyclopentane	<SRL	U	1	0.25	0.17
2,4-Dimethylpentane	<SRL	U	1	0.21	0.14
Benzene	<SRL	U	1	0.25	0.17
Cyclohexane	<SRL	U	1	0.25	0.17
2-Methylhexane	<SRL	U	1	0.21	0.14
2,3-Dimethylpentane	<SRL	U	1	0.21	0.14
3-Methylhexane	<SRL	U	1	0.21	0.14
2,2,4-Trimethylpentane	<SRL	U	1	0.19	0.13
n-Heptane	<SRL	U	1	0.21	0.14
Methylcyclohexane	<SRL	U	1	0.21	0.14
2,3,4-Trimethylpentane	<SRL	U	1	0.19	0.13



Atmospheric Analysis & Consulting, Inc.

Laboratory Analysis Report


CLIENT : SLR International Corporation
PROJECT NO : 220526 Rev 1
MATRIX : AIR
UNITS : ppb (v/v)

DATE RECEIVED : 03/15/2022
DATE REPORTED : 03/16/2022

HYDROCARBONS (C2-C12) SPECIATED

<i>Client ID</i>	NUI DUP			Sample Reporting Limit (SRL) (MRLxDFs)	Method Reporting Limit (MRL)
<i>AAC ID</i>	220526-28954				
<i>Date Sampled</i>	03/12/2022				
<i>Date Analyzed</i>	03/15/2022				
<i>Car Dilution Factor</i>	1.49				
	Result	Qualifier	Analysis DF		
Toluene	<SRL	U	1	0.21	0.14
2-Methylheptane	<SRL	U	1	0.19	0.13
3-Methylheptane	<SRL	U	1	0.19	0.13
n-Octane	<SRL	U	1	0.19	0.13
Ethylbenzene	<SRL	U	1	0.19	0.13
m/p-Xylenes	<SRL	U	1	0.19	0.13
Styrene	<SRL	U	1	0.19	0.13
o-Xylene	<SRL	U	1	0.19	0.13
Nonane	<SRL	U	1	0.17	0.11
Isopropylbenzene	<SRL	U	1	0.17	0.11
n-Propylbenzene	<SRL	U	1	0.17	0.11
m-Ethyltoluene	<SRL	U	1	0.17	0.11
p-Ethyltoluene	<SRL	U	1	0.17	0.11
1,3,5-Trimethylbenzene	<SRL	U	1	0.17	0.11
o-Ethyltoluene	<SRL	U	1	0.17	0.11
1,2,4-Trimethylbenzene	<SRL	U	1	0.17	0.11
n-Decane	<SRL	U	1	0.15	0.10
1,2,3-Trimethylbenzene	<SRL	U	1	0.17	0.11
m-Diethylbenzene	<SRL	U	1	0.15	0.10
p-Diethylbenzene	<SRL	U	1	0.15	0.10
n-Undecane	<SRL	U	1	0.14	0.09
n-Dodecane	<SRL	U	1	0.12	0.08

U - Compound was analyzed for, but was not detected at or above the SRL.



Atmospheric Analysis & Consulting, Inc.

Quality Control/Quality Assurance Report PAMS Calibration Verification Analysis

Initial Calibration Date : 02/11/2022
Standard ID : MS1-020922-01

Instrument ID : MS01
Analysis Date : 03/14/2022
Analyst : RB

Continuing Calibration Verification

Propane	xRF	Daily RF	RPD*
	698	691	1.04


* Must be <10%

Laboratory Control Spike (LCS) / Laboratory Control Spike Duplicate (LCSD)

Propane	Sample Conc. (ppbC)	Spike Added (ppbC)	Recovery (ppbC)		% Recovery**		RPD***
			LCS	LCSD	LCS	LCSD	
			0.00	4.24	4.19	4.18	

** Must be 80-120%

*** Must be <25%



Atmospheric Analysis & Consulting, Inc.


Quality Control/Quality Assurance Report PAMS Method Blank Analysis

Matrix : Air
Units : ppbC

Instrument ID : MS01
Analysis Date : 03/14/2022
Analyst : RB

Analyte	Result	PQL
Ethylene	<PQL	1.0
Acetylene	<PQL	1.0
Ethane	<PQL	1.0
Propylene	<PQL	1.0
Propane	<PQL	1.0
Isobutane	<PQL	1.0
1-Butene	<PQL	1.0
n-Butane	<PQL	1.0
trans-2-Butene	<PQL	1.0
cis-2-Butene	<PQL	1.0
Isopentane	<PQL	1.0
1-Pentene	<PQL	1.0
n-Pentane	<PQL	1.0
Isoprene	<PQL	1.0
trans-2-Pentene	<PQL	1.0
cis-2-Pentene	<PQL	1.0
2,2-Dimethylbutane	<PQL	1.0
Cyclopentane	<PQL	1.0
2,3-Dimethylbutane	<PQL	1.0
2-Methylpentane	<PQL	1.0
3-Methylpentane	<PQL	1.0
1-Hexene	<PQL	1.0
n-Hexane	<PQL	1.0
Methylcyclopentane	<PQL	1.0
2,4-Dimethylpentane	<PQL	1.0
Benzene	<PQL	1.0
Cyclohexane	<PQL	1.0
2-Methylhexane	<PQL	1.0
2,3-Dimethylpentane	<PQL	1.0
3-Methylhexane	<PQL	1.0
2,2,4-Trimethylpentane	<PQL	1.0
n-Heptane	<PQL	1.0
Methylcyclohexane	<PQL	1.0
2,3,4-Trimethylpentane	<PQL	1.0

Analyte	Result	PQL
Toluene	<PQL	1.0
2-Methylheptane	<PQL	1.0
3-Methylheptane	<PQL	1.0
n-Octane	<PQL	1.0
Ethylbenzene	<PQL	1.0
m/p-Xylenes	<PQL	1.0
Styrene	<PQL	1.0
o-Xylene	<PQL	1.0
Nonane	<PQL	1.0
Isopropylbenzene	<PQL	1.0
n-Propylbenzene	<PQL	1.0
m-Ethyltoluene	<PQL	1.0
p-Ethyltoluene	<PQL	1.0
1,3,5-Trimethylbenzene	<PQL	1.0
o-Ethyltoluene	<PQL	1.0
1,2,4-Trimethylbenzene	<PQL	1.0
n-Decane	<PQL	1.0
1,2,3-Trimethylbenzene	<PQL	1.0
m-Diethylbenzene	<PQL	1.0
p-Diethylbenzene	<PQL	1.0
n-Undecane	<PQL	1.0
n-Dodecane	<PQL	1.0
TNMHC (ppbC)	<PQL	20



Atmospheric Analysis & Consulting, Inc.

Quality Control/Quality Assurance Report PAMS Duplicate Analysis

AAC ID : 220507-28820
 Matrix : Air
 Units : ppbC

Instrument ID : MS01
 Analysis Date : 03/14/2022
 Analyst : RB

Analyte	Sample Analysis	Sample Duplicate	RPD
Ethylene	<PQL	<PQL	NA
Acetylene	<PQL	<PQL	NA
Ethane	4.69	5.25	11
Propylene	<PQL	<PQL	NA
Propane	2.79	2.96	5.9
Isobutane	<PQL	<PQL	NA
1-Butene	<PQL	<PQL	NA
n-Butane	<PQL	<PQL	NA
trans-2-Butene	<PQL	<PQL	NA
cis-2-Butene	<PQL	<PQL	NA
Isopentane	<PQL	<PQL	NA
1-Pentene	<PQL	<PQL	NA
n-Pentane	<PQL	<PQL	NA
Isoprene	<PQL	<PQL	NA
trans-2-Pentene	<PQL	<PQL	NA
cis-2-Pentene	<PQL	<PQL	NA
2,2-Dimethylbutane	<PQL	<PQL	NA
Cyclopentane	<PQL	<PQL	NA
2,3-Dimethylbutane	<PQL	<PQL	NA
2-Methylpentane	<PQL	<PQL	NA
3-Methylpentane	<PQL	<PQL	NA
1-Hexene	<PQL	<PQL	NA
n-Hexane	<PQL	<PQL	NA
Methylcyclopentane	<PQL	<PQL	NA
2,4-Dimethylpentane	<PQL	<PQL	NA
Benzene	<PQL	<PQL	NA
Cyclohexane	<PQL	<PQL	NA
2-Methylhexane	<PQL	<PQL	NA
2,3-Dimethylpentane	<PQL	<PQL	NA
3-Methylhexane	<PQL	<PQL	NA
2,2,4-Trimethylpentane	<PQL	<PQL	NA
n-Heptane	<PQL	<PQL	NA
Methylcyclohexane	<PQL	<PQL	NA
2,3,4-Trimethylpentane	<PQL	<PQL	NA
Toluene	<PQL	<PQL	NA
2-Methylheptane	<PQL	<PQL	NA
3-Methylheptane	<PQL	<PQL	NA
n-Octane	<PQL	<PQL	NA
Ethylbenzene	<PQL	<PQL	NA
m/p-Xylenes	<PQL	<PQL	NA
Styrene	<PQL	<PQL	NA
o-Xylene	<PQL	<PQL	NA
Nonane	<PQL	<PQL	NA
Isopropylbenzene	<PQL	<PQL	NA
n-Propylbenzene	<PQL	<PQL	NA
m-Ethyltoluene	<PQL	<PQL	NA
p-Ethyltoluene	<PQL	<PQL	NA

Analyte	Sample Analysis	Sample Duplicate	RPD
1,3,5-Trimethylbenzene	<PQL	<PQL	NA
o-Ethyltoluene	<PQL	<PQL	NA
1,2,4-Trimethylbenzene	<PQL	<PQL	NA
n-Decane	<PQL	<PQL	NA
1,2,3-Trimethylbenzene	<PQL	<PQL	NA
m-Diethylbenzene	<PQL	<PQL	NA
p-Diethylbenzene	<PQL	<PQL	NA
n-Undecane	<PQL	<PQL	NA
n-Dodecane	<PQL	<PQL	NA
Total PAMS (ppbC)	7.47	8.20	9.3
TNMHC (ppbC)	30.6	34.9	13